



California Medical Device Recall Information



Recall Name

Hamilton Medical Recalls the G5 Ventilator Due to a Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
11/13/15	Hamilton- G5 Ventilator <ul style="list-style-type: none">Software Version V2.00Software Version V2.31	Hamilton Medical, Inc. Reno, NV	<i>Potential for ventilator to fail when operator activates suctioning maneuver.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Catalog Number: 1590001 List of Suspect S/Ns	CA , nationwide	Distributed between: March 2007 to March 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/MedWatch/SafetyInformation/SafetyAlertsforHumanMedicalProducts/ucm471725.htm>